

湖北省计量测试技术研究院

Hubei Institute of Measurement and Testing Technology

校准证书

Calibration Certificate

证书编号: 2022DW034000337
Certificate No.

委托方 Client	湖南三一智慧新能源设计有限公司
委托方地址 Address	湖南省长沙市长沙县三一路三一重工产业园2号办公楼
器具名称 Name of instrument	接地电阻测试仪
制造厂商 Manufacturer	武汉国电华美电气设备有限公司
型号/规格 Type/Specification	HM2571
器具编号 Serial No.	02203008

湖北省计量测试技术研究院
证书骑缝章



批准人
Approved by 耿睿

核验员
Checked by 普夏

校准员
Calibrated by 王强

样品接收日期 Date of Application	2022	年	06	月	16	日
校准日期 Date of Calibration	2022	年	06	月	23	日
签发日期 Date of Issue	2022	年	06	月	23	日



国家法定计量检定机构计量授权证书号: (国)法计(2017)01028号

地址: 湖北省武汉市东湖新技术开发区茅店山中路二号(总部)
Add: No.2, Maodianshanzhong Road, East Lake High-tech Development Zone, Wuhan, Hubei
网址 (Web site): <http://www.himtt.net>

邮编 (Post Code): 430223
电话 (Tel): 027-81925136
传真 (Fax): 027-81925137

● 本院是政府计量行政管理部门依法设立的国家法定计量检定机构

This laboratory is a legal metrological verification institution established by the government metrological administrative department according to law.

● 本次校准的技术依据(代号、名称)

Reference documents for the Calibration (Code, Name)

参照 JJG 366-2004 《接地电阻表检定规程》

● 本次校准所使用的主要计量标准器具

Main standards of measurement used in the Calibration

名称 Name	型号/编号 Type/Serial No.	测量范围 Measuring Range	不确定度/准确度等级/ 最大允许误差 Uncertainty/Accuracy Class/ Maximum Permissible Error	证书编号/ 溯源机构 Certificate No / Traceability Agency	有效期限 Due date
接地电阻表检定装置	JDB-2 96115	(0.001-10000) Ω	0.05级	2022DW02250053 湖北省计量测试 技术研究院	2023-03-09

● 校准地点及环境条件

Environmental condition and Place for the Calibration

地点: 茅店山B211

Site

温度: 20.0 $^{\circ}\text{C}$

Temperature

相对湿度: 40 %

R.H.

其它: /

Others

气压: / kPa

Pressure

原始记录编号: 2022DW034000337

Record No.

● 备注 /

Note

- 1、本院所出具的量值(校准结果), 均可溯源至国家计量基准和(或)国际单位制(SI)。
All data issued by this laboratory are traceable to national primary standards and international system of units(SI).
- 2、校准结果, 仅对受检样品的本次校准有效。校准证书封面未加盖校准专用章无效。
It's Effect That Results of This Report Relate Only To The Sample(s) Calibrated. It's Invalid That The Certificate Cannot Be Stamped.
- 3、未经本院书面批准, 不得复制(全文复制除外)。
Without the written approval of the court, it is not allowed to copy (except full-text copy)
- 4、“委托方”、“委托方联络信息”由委托方提供, “制造厂”、“型号规格”、“器具编号”为仪器上标注。
The information Client and Contact Information are provided by client, and the Manufacturer, Type /Specification, Serial No. are marked on the items.

校准结果

Results of Calibration

量程: 200 Ω		量程: 20 Ω		量程: 2 Ω	
标准值(Ω)	显示值 (Ω)	标准值(Ω)	显示值 (Ω)	标准值(Ω)	显示值 (Ω)
10.00	10.0	1.000	1.00	0.100	0.100
20.00	20.0	2.000	1.99	0.200	0.200
30.00	29.9	3.000	2.99	0.300	0.299
40.00	39.9	4.000	3.99	0.400	0.398
50.00	49.9	5.000	4.99	0.500	0.498
60.00	59.8	6.000	5.99	0.600	0.597
70.00	69.8	7.000	6.99	0.700	0.697
80.00	79.8	8.000	7.99	0.800	0.797
90.00	89.7	9.000	8.99	0.900	0.898
100.0	99.7	10.00	9.98	1.000	0.999
200.0	199.1	20.00	19.95	2.000	1.997
$U_{rel} (k=2)$	1%	$U_{rel} (k=2)$	1%	$U_{rel} (k=2)$	1%

以下空白



